



ULTRA-TURRAX® Tube Drive

/// Data Sheet

Unique and universal dispersing, stirring, homogenizing and grinding system with sealable disposable sample tubes. Protection and security for: Infectious sample materials, toxic substances, high-odor substances under defined conditions (time, energy, volume).

For research purposes only! Not suitable for diagnostic use (in accordance with the IVD Directive).

- Disperse, stir, homogenize and grind using a single drive unit
- No possibility of cross-contamination









designed for scientists

- Sealable disposable sample tubes
- No cleaning required
- High level of user safety
- Suitable for individual use and use in series
- Gamma-sterilized tubes
- Tubes with pierceable membrane covers
- Tubes with 2 15 ml and 15 50 ml
- Reproducible tests
- Patented

Application areas: Human medicine, pathology, veterinary medicine, animal hygiene institutes, Foodstuffs testing laboratories, Toxicology, Medical research, Pharmaceutical research, Biological research, Tumour biology, Chemistry, Cosmetics





designed for scientists

Technical Data

Motor rating input [W]	20	
Motor rating output [W]	17	
Viscosity max. [mPas]	5000	
Speed range [rpm]	300 - 6000	
Speed deviation [%]	10	
Speed adjustment	stepless	
Speed display	scale	
Noise without element [dB(A)]	50	
Process type	batch	
Dimensions (W x H x D) [mm]	100 x 40 x 160	
Weight [kg]	0.75	
Permissible ambient temperature [°C]	5 - 40	
Permissible relative humidity [%]	80	
Protection class according to DIN EN 60529	IP 20	
Voltage [V]	100 - 240	
Frequency [Hz]	50/60	
Power input [W]	20	
DC Voltage [V=]	24	
Current consumption [mA]	800	





Wolf Laboratories Limited

www.wolflabs.co.uk

Tel: 01759 301142

Fax:01759 301143

sales@wolflabs.co.uk







Use the above details to contact us if this literature doesn't answer all your questions.

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.





