



*USB 2.0 C-mount and Eyepiece
Microscope Camera*

General specifications	USB output	Yes
	C-Mount	Yes
	Dimensions (mm)	58 x 58 x 36.4
	Weight (kg)	0.18

Sensor	Camera resolution (n° of pixels: W x H)	2048x1536
	Color / Monochrome	Color
	Sensor Size	1/2.8"
	Sensor Technology	CMOS
	Sensor Type	APTINA CMOS
	Image Format	4/3
	Pixel Size (mm)	2.5x2.5
	Acquisition technology	Rolling shutter
	G sensitivity	600mV at 1/30s
	Dark Signal	0.15mV at 1/30s
	ADC conversion	8 Bit
	Color Depth	1, 4, 8, 24 Bit
	Exposure time	0.244 msec - 2 sec
	Binning	1x1; 2x2; 3x3
	IR filter	Yes
IR filter range (nm)	380-650 (IR CUT)	

USB	PC Camera resolution (MP)	3.1
	USB standard	USB2.0
	Frame rate full resolution (fps)	10.5@2048x1536
	Frame rate other resolution (fps)	15@1920x1080

Camera Power Supply	Connector	USB
	Output voltage	PC USB

Camera accessories	Optical adapter (for 23mm eyepiece tube)	0.5x
	Adapter ring	Yes
	Diameter of adapter ring (mm)	30
	Diameter of adapter ring (mm)	30.5
	Calibration slide	Yes
	USB cable (1.8m)	Yes

Environment	Working Temperature	-10°C up to +50°C
	Storage Temperature	-20°C up to +60°C
	Relative Humidity	30% to 80%
	Storage Humidity	10% to 60%

Operating System	OS	Windows 7
	OS	Windows 8
	OS	Windows 10 (32Bit)
	OS	Windows 10 (64Bit)
	OS	Windows 11
	OS	Osx (MAC Osx) starting from 10.7.5
	OS	Linux

Available Software	PC SW	PROVIEW
	PC SW	LITEVIEW

PC requirements	PC	CPU: Intel Core2 - 2.8GHz or higher
	PC	RAM: 2GB or more
	PC	USB 2.0



Wolflabs

Wolf Laboratories Limited

www.wolflabs.co.uk

Tel: 01759 301142

Fax: 01759 301143

sales@wolflabs.co.uk



Use the above details to contact us if this literature doesn't answer all your questions.

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.

