

Specification Sheet

Specification Number: 8055H-1000S
Revision: 09
Date Issued: 29-Jun-21

1. Part Number: 8055H-1000S

2. Photograph:





3. Product Information:

Description: SteriWare® ViscoThief - 1000mm

Use: Single use, disposable liquid sampler (for high viscosity liquids)

Material of Construction: High Density Polyethylene (HDPE) - virgin

Conforms to FDA 21 CFR 177.1520
Conforms to EU Regulations 10/2011
Conforms to EC Regulations 1935/2004

• Conforms to REACH regulation 1907/2006

• BSE/TSE free

O-ring Seal – virgin Silicone Rubber

• Conforms to FDA CFR 177.2600

Method of Construction: Ultrasonic welding

Nominal Length

Fully closed: 991mm
Fully extended: 1818mm

Diameter of Outer Tube: 21mm

Diameter of hole in tip: 10mm

Width of Handle: 103mm

Nominal Weight of Sampler: 158g

Assembly & Primary Packing Environment: Class 100,000 Cleanroom Individually Bagged? Yes (heat sealed PE/PET bag)

Method of Sterilisation: Gamma Irradiation (25kGy minimum)

Number of Samplers per Box: 20

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture





Specification Sheet

Specification Number:8055H-1000SRevision:09Date Issued:29-Jun-21

4. Change Control: Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

5. Country of Origin: Made and packed in the UK.

6. Repeated Use: This is a single-use, disposable sampler. Repeated use will lead to

failure.

7. Approval: Signed:

Name: Kuldeep Flora

Quality Manager

Date: 29-Jun-21

Position:





Wolf Laboratories Limited

www.wolflabs.co.uk

Tel: 01759 301142

Fax:01759 301143

sales@wolflabs.co.uk







Use the above details to contact us if this literature doesn't answer all your questions.

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.





