

### **Specification Sheet**

Specification Number: 8088R-10S
Revision: 03
Date Issued: 11-Dec-19

**1. Part Number**: 8088R-10S

### 2. Photograph:



#### 3. Product Information:

Description: SteriWare Powder Spoon Red - 10ml - Sterile

Use: Single use, disposable spoon

Material of Construction: High Impact Polystyrene (HIPS) – rigid, red

FDA acceptable CFR 177.1640EU Regulations 10/2011 compliant

• EC Regulations 1935/2004 compliant

BSE/TSE free

Method of Construction: Single piece, injection moulded

Nominal Bowl Volume: 10ml
Length: 185mm
Bowl Length: 48mm
Bowl Width: 35mm
Nominal Weight: 6g

Moulding & Primary Packing Environment: Class 100,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)
Method of Sterilisation: Gamma Irradiation (25-45kGy)

Number of Samplers per Box: 100

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This product has not been ATEX tested





# **Specification Sheet**

Specification Number:8088R-10SRevision:03Date Issued:11-Dec-19

**4. Change Control:** Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

**5. Country of Origin:** Made and packed in the UK.

**6. Repeated Use:** This is a single-use, disposable sampler. Repeated use will lead to

failure.

7. Approval: Signed: Name: Kuldeep Flora

Position: Quality Manager

Date: 11-Dec-19





Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.

## www.wolflabs.co.uk

Tel: 01759 301142

Fax: 01759 301143

sales@wolflabs.co.uk

Please contact us if this literature doesn't answer all your questions.